 <b>Botswana Medicines Regulatory Authority</b>	Page 1 of 20
	<b>Document type:</b> Guideline
	<b>Title:</b> Guideline for Medical Devices Classification
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023

## Botswana Medicines Regulatory Authority




Approved By: \_\_\_\_\_

**Mr. Bathusi Kgosietsile**  
**Director - Product Evaluations**  
**and Registration**

\_\_\_\_\_


**Date of approval (DD/MM/YY)**



 <b>Botswana Medicines Regulatory Authority</b>	Page 3 of 20
	<b>Document type:</b> Guideline
	<b>Title:</b> Guideline for Medical Devices Classification
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023

## Table of Contents

Revision status sheet.....	2
1. Purpose.....	4
2. Scope.....	4
3. Definitions and Abbreviations .....	4
3.1 Definitions .....	4
3.2 Abbreviations .....	7
4. General Principles .....	7
5. Risk Classification System for Medical Devices .....	8
6. Classification Rules for Medical Devices.....	10
7. Rationale for inclusion of Additional Rules .....	20

 <b>Botswana Medicines Regulatory Authority</b>	Page 4 of 20
	<b>Document type:</b> Guideline
	<b>Title:</b> Guideline for Medical Devices Classification
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023

## 1. Purpose

The intention and purpose of this guideline is to provide a series of principles and rules that allow a medical device to be assigned to one of four classes based on its intended use, and thereby assist a manufacturer to allocate its medical device to an appropriate class using a set of classification rules.

## 2. Scope

This guideline applies to all products that fall within the definition of the term 'medical device', excluding in vitro diagnostic medical devices, for which a separate classification guideline is available.

## 3. Definitions and Abbreviations

### 3.1 Definitions

The following definitions shall apply:

#### 3.1.1 Accessory to a Medical Device

An article intended specifically by its manufacturer to be used together a particular medical device to enable or assist that device to be used in accordance with its intended use.

#### 3.1.2 Active Medical Device

Any medical device, operation of which depends on a source of electrical energy or any source of power other than that directly generated by the human body or gravity and which acts by converting this energy. Medical devices intended to transmit energy, substances or other elements between an active medical device and the patient, without any significant change, are not considered to be active medical devices. Standalone software is considered to be an active medical device.

#### 3.1.3 Active Therapeutic Device

Any active medical device, whether used alone or in combination with other medical devices, to supply information for detecting, diagnosing, monitoring or treating physiological conditions, states of health, illnesses or congenital deformities.


#### 3.1.4 Active Device Intended for Diagnosis

Any active medical device, whether used alone or in combination with other medical devices, to supply information for detecting, diagnosing, monitoring or treating physiological conditions, states of health, illnesses or congenital deformities.

**3.1.5 Body orifice:** Any natural opening in the body, as well as the external surface of the eyeball, or any permanent artificial opening, such as a stoma or permanent tracheotomy

#### 3.1.5 Central Circulatory System

The major internal blood vessels including the following: pulmonary veins, pulmonary arteries, cardiac veins, coronary arteries, carotid arteries (common, internal and external), cerebral arteries, brachiocephalic artery, aorta (includes all segments of the aorta), inferior and superior vena cava and common iliac arteries.

 <b>Botswana Medicines Regulatory Authority</b>	Page 5 of 20
	<b>Document type:</b> Guideline
	<b>Title:</b> Guideline for Medical Devices Classification
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023

### 3.1.6 **Central nervous system:**

The brain, meninges, and spinal cord.

### 3.1.7 **Label**

In relation to a health product or an active ingredient, means any written, printed or graphic representation that appears on or is attached to the health product or active ingredient or any part of its packaging, and includes any informational sheet or leaflet that accompanies the health product or active ingredient when it is being supplied.

### 3.1.8 **Cleaning**

Removal of contamination from an item to the extent necessary for its further processing and its intended subsequent use.

### 3.1.9 **Disinfection**

Reduction of the number of viable microorganisms on a product to a level previously specified as appropriate for its intended further handling or use.

### 3.1.9 **Duration of use**

**Transient:** Normally intended for continuous use for less than 60 minutes.

**Short term:** Normally intended for continuous use for between 60 minutes and 30 days.

**Long term:** Normally intended for continuous use for more than 30 days.

**NOTE:** continuous use means:

- The entire duration of use of the same device without regard to temporary interruption of use during a procedure or, temporary removal for purposes such as cleaning or disinfection of the device.
- The accumulated use of a device that is intended by the manufacturer to be replaced immediately with another of the same type.

### 3.1.9 **Harm**

Physical injury or damage to the health of people or damage to property or the environment.

### 3.1.9 **Hazard**

Potential source of harm.


### 3.1.9.4 **Implantable device:** Any device, including those that are partially or wholly absorbed, which is intended: -

- i. to be totally introduced into the human body or,
  - ii. to replace an epithelial surface or the surface of the eye,
- by surgical intervention which is intended to remain in place after the procedure. Any device intended to be partially introduced into the human body through surgical intervention and intended to remain in place after the procedure for at least 30 days is also considered an implantable device.

### 3.1.9 **Intended use / Intended purpose:**

The objective intent of the manufacturer regarding the use of a product, process or service as reflected in the specifications, instructions and information provided by the manufacturer.

**This document is property of the Botswana Medicines Regulatory Authority (BOMRA). It is strictly confidential and may on no account be reproduced, copied or divulged to any third party without prior authorization by BOMRA Management.**

 <b>Botswana Medicines Regulatory Authority</b>	Page 6 of 20
	<b>Document type:</b> Guideline
	<b>Title:</b> Guideline for Medical Devices Classification
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023

3.1.9 **Invasive device:** A device, which, in whole or in part, penetrates inside the body, either through a body orifice or through the surface of the body.

3.1.10 **Lay Person:**

Individual that does not have formal training in a relevant field or discipline.

3.1.11 **Life supporting or life sustaining**

A device that is essential to, or that yields information that is essential to, the restoration or continuation of a bodily function important to the continuation of human life.

3.1.12 **Medical device**

It means any instrument, apparatus, implement, machine, appliance, implant, in vitro reagent or calibrator, software, material or other similar or related article -

- a) intended by the manufacturer to be used, alone or in combination, for humans or animals for-
  - i. diagnosis, prevention, monitoring, treatment or alleviation of disease;
  - ii. diagnosis, monitoring, treatment, alleviation of or compensation for an injury;
  - iii. investigation, replacement, modification or support of the anatomy or of a physiological process;
  - iv. supporting or sustaining life;
  - v. control of conception;
  - vi. cleaning, disinfection or sterilization of medical devices; or
  - vii. providing information for medical or diagnostic purpose by means of in vitro examination of specimens derived from the human body; and
- b) which do not achieve its primary intended action in or on human or animal body by pharmacological, immunological or metabolic means but which may be assisted in its intended function by such means

3.1.14 **Reusable medical device**

Means a device intended for repeated use either on the same or different patients, with appropriate decontamination and other reprocessing between uses.

3.1.15 **Reusable surgical instrument**

Instrument intended for surgical use by cutting, drilling, sawing, scratching, scraping, clamping, retracting, clipping or similar surgical procedures, without connection to any active medical device and which are intended by the manufacturer to be reused after appropriate procedures for cleaning and/or sterilisation have been carried out.


3.1.16 **Risk**

Combination of the probability of occurrence of harm and the severity of that harm.

3.1.18 **Sterilisation**

Validated process used to render product free from viable microorganisms.

NOTE: In a sterilisation process, the nature of microbial inactivation is exponential and thus the survival of a microorganism on an individual item can be expressed in terms of probability. While this probability can be reduced to a very low number, it can never be reduced to zero.

 <b>Botswana Medicines Regulatory Authority</b>	Page 7 of 20
	<b>Document type:</b> Guideline
	<b>Title:</b> Guideline for Medical Devices Classification
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023

### 3.1.9.3 Surgically invasive device:

- (a) An invasive device which penetrates inside the body through the surface of the body, with the aid or in the context of a surgical operation.
- (b) A medical device which produces penetration other than through a body orifice.

### 3.1.20 Vital physiological process

Means a process that is necessary to sustain life, the indicators of which may include any one or more of the following:

- i. respiration;
- ii. heart rate;
- iii. cerebral function;
- iv. blood gases;
- v. blood pressure;
- vi. body temperature.


## 3.2 Abbreviations

The following abbreviations shall apply:

- 3.2.1 **CE**- European Conformity
- 3.2.2 **IMDRF**- International Medical Device Regulators Forum.
- 3.2.3 **IVD** - In Vitro Diagnostic
- 3.2.4 **GHTF**- The Global Harmonization Task Force
- 3.2.5 **WHO** - World Health Organization.

## 4. General Principles

Manufacturers of medical devices are subject to regulatory controls overseen by Authority. The Authority specifies procedures to be followed by manufacturers during the design, manufacture, and marketing of each device, and describes the manner in which a manufacturer should demonstrate conformity to such procedures. It is widely accepted that oversight of these procedures by the Authority should increase in line with the potential of a medical device to cause harm to a patient or user (i.e. The hazard it presents), and to do so without placing an unwarranted increase in the cost of regulatory compliance for either the manufacturer or Authority, or delay market entry. In practice, this is best achieved by assigning every medical device into one of four groups – or ‘classes’ - by applying the classification rules described in this guidance document, and specify in separate guidance the different conformity assessment procedures that should apply to each group of devices described in this guidance. Customarily, a classification system of this type is referred to as a ‘risk-based classification scheme’ but this is a misnomer since the rules take account only of the hazard presented by

 <b>Botswana Medicines Regulatory Authority</b>	Page 8 of 20
	<b>Document type:</b> Guideline
	<b>Title:</b> Guideline for Medical Devices Classification
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023

a particular device and not the probability harm will occur document, and specify in separate requirements at the conformity assessment stage rather than modifying the classification rules. The Authority takes account of the probability that harm will occur by modifying the evidence requirements at the conformity assessment stage rather than modifying the classification rules. Probability of harm is influenced by factors such as whether:

- A. the technology is regarded as mature;
- B. the device type is the source of many adverse event reports;
- C. the device's manufacturer has a long experience of the device and the technologies it embodies;
- D. the device user is a lay person.


The hazard presented by a particular medical device depends substantially on its intended use and the technology it utilises. Consequently, the classification rules stipulated in section 6.0 of this guidance document take factors into account such as, whether the device:

- a) Is life supporting or sustaining;
- b) is invasive and if so, to what extent and for how long;
- c) incorporates medicinal products, or human/animal tissues/cells;
- d) is an active medical device;
- e) delivers medicinal products, energy or radiation;
- f) could modify blood or other body fluids;
- g) is used in combination with another medical device.

## 5. Risk Classification System for Medical Devices

### 5.1 Structure of the Classification Rules


- a) The device classification system consists of four classes where Class A represents the lowest risk and Class D the highest.
- b) The determination of class is based on rules derived from the potential of a medical device to cause harm to a patient or user (i.e. the hazard it presents) and thereby on its intended use and the technology/ies it utilises.
- c) These rules allow a manufacturer to readily identify the class of its particular medical device subject., where appropriate, to the Authority resolving any matters of interpretation.
- d) The rules should be capable of accommodating future technological developments.
- e) The manufacturer should document its justification for placing its product into a particular class, including the resolution of any matters of interpretation where it has asked the Authority for a ruling.
- f) An accessory to a medical device may be classified separately using the classification rules in this guidance document

 <b>Botswana Medicines Regulatory Authority</b>	Page 9 of 20
	<b>Document type:</b> Guideline
	<b>Title:</b> Guideline for Medical Devices Classification
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023

- g) If, based on the manufacturer's intended use, two or more classification rules apply to the device, the device is allocated the highest level of classification indicated.
- h) Where one medical device is intended to be used together with a different medical device, that may or may not be from the same manufacturer, (e.g. a pulse oximeter and a replaceable sensor sourced from a different manufacturer, or a general purpose syringe and a syringe driver), the classification rules should apply separately to each of the devices.
- i) Classification of an assemblage of medical devices that individually comply with all regulatory requirements depends on the manufacturer's purpose in packaging and marketing such a combination of separate devices. For example:
- i. If the combination results in a product that is intended by the manufacturer to meet a purpose different from that of the individual medical devices that make it up, the combination is a new medical device in its own right and should be classified according to the new intended use.
  - ii. If the combination is for the convenience of the user but does not change the intended uses of the individual medical devices that make it up (e.g. a customised kit that provides all the devices necessary to carry out a particular surgical procedure), the classification allocated to the assemblage for the purpose of a Declaration of Conformity is at the level of the highest classified device included within it.
- j) Classification of an assemblage of medical devices where one or more of the medical devices that is in the assemblage has yet to comply with all the relevant regulatory requirements, should be for the combination as a whole according to its intended use.
- k) While most software is incorporated into the medical device itself, some is not. Provided such standalone software falls within the scope of the definition for a 'medical device', it is deemed to be an active device and should be classified as follows:
- a. Where it drives or influences the use of a separate medical device, it should be classified according to the intended use of the combination.
  - b. Where it is independent of any other medical device, it is classified in its own right using the rules in Section 6.0 of this document.
- l) The historical experience of the Authority may require a particular type of medical device to be allotted/allocated a different classification from that assigned through the application of these classification rules.

## 5.2 Diagrammatic Representation of the Classification System

Table below indicate the four risk classes of medical devices

 <b>Botswana Medicines Regulatory Authority</b>	Page 10 of 20
	<b>Document type:</b> Guideline
	<b>Title: Guideline for Medical Devices Classification</b>
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023

Risk Class	Risk Level	Examples
<b>A</b>	Low Risk/ Hazard	Bandages / tongue depressors Stethoscope, Examination gloves
<b>B</b>	Low to Moderate Risk	Infusion set, Suction Equipment Hypodermic Needles / suction equipment
<b>C</b>	Moderate Risk to high Risk	X-ray, Ventilator Lung ventilator / bone fixation plate
<b>D</b>	High Risk	Heart Valve, pacemaker


The examples of medical devices provided in table above are for illustration only and a manufacturer of such a device should not rely on it appearing as an example but should instead make an independent decision on classification taking account of its particular design and intended use

## 6. Classification Rule for Medical Devices


The actual classification of each device depends on the claims made by the manufacturer for its intended use and the technology/ies it utilises. As an aid to interpreting the purpose of each rule, illustrative examples of medical devices that should conform to the rule have been provided in the table below. However, it must be emphasised that a manufacturer of such a device should not rely on it appearing as an example but should instead make an independent decision on classification taking account of its particular design and intended use.

### 6.1. Non – Invasive Devices

Rule	Illustrative Examples
<b>Rule 1.</b> All non-invasive devices which come into contact with injured skin:	Devices covered by this rule are extremely claim sensitive.
- are in Class A if they are intended to be used as a mechanical barrier, for compression or for absorption of exudates only, i.e. wound that heal by primary intent;	<u>Examples:</u> bandages; cotton wool
- are in Class B if they are intended to be used principally with wounds which have breached the dermis, including devices principally intended to manage the microenvironment of a wound.	<u>Example:</u> non-medicated impregnated gauze dressings.
<b>unless</b> they are intended to be used principally with wounds which have	Devices used to treat wounds where the subcutaneous tissue is at least partially exposed

 <b>Botswana Medicines Regulatory Authority</b>	Page 11 of 20
	<b>Document type:</b> Guideline
	<b>Title:</b> Guideline for Medical Devices Classification
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023


breached the dermis and can only heal by secondary intent, in which case they are in Class C	and the edges of the wound are not sufficiently close to be pulled together. To close the wound, new tissue must be formed within the wound prior to external closure. The device manufacturer claims that they promote healing through physical methods other than 'primary intent'.  <u>Examples:</u> dressings for chronic ulcerated wounds; dressings for severe burns.
<b>Rule 2(i).</b> All non-invasive devices intended for channelling or storing <ul style="list-style-type: none"> <li>liquids, or</li> <li>gases</li> </ul> for the purpose of eventual infusion, administration or introduction into the body are in Class A	Such devices are 'indirectly invasive' in that they channel or store liquids that will eventually be delivered into the body.  <u>Examples:</u> administration sets for gravity infusion; syringes without needles.
<b>unless</b> they may be connected to an active medical device in Class B or a higher class, in which case they are Class B	<u>Examples:</u> syringes and administration sets for infusion pumps; anaesthesia breathing circuits.  <b>NOTE:</b> "Connection" to an active device covers those circumstances where the safety and performance of the active device is influenced by the non-active device and vice versa
<b>Rule 2(ii).</b> All non-invasive devices intended to be used for <ul style="list-style-type: none"> <li>channelling blood, or</li> <li>storing or channelling other body liquids, or</li> <li>storing organs, parts of organs or body tissues,</li> </ul> for the purpose of eventual infusion, administration or introduction into the body are Class B.	<u>Examples:</u> tubes used for blood transfusion, organ storage containers, tissue and cell containers.
<b>unless</b> they are blood bags, in which case they are Class C.	<u>Example:</u> Blood bags that do not incorporate an anti-coagulant, blood component storage bags.

 <b>Botswana Medicines Regulatory Authority</b>	Page 12 of 20
	<b>Document type:</b> Guideline
	<b>Title: Guideline for Medical Devices Classification</b>
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023

<p><b>Rule 3.</b> All non-invasive devices intended for modifying the biological or chemical composition of</p> <ul style="list-style-type: none"> <li>• blood,</li> <li>• other body liquids, or</li> <li>• other liquids,</li> </ul> <p>intended for infusion into the body are in Class C,</p>	<p>Such devices are ‘indirectly invasive’ in that they treat or modify substances that will eventually be delivered into the body. They are normally used in conjunction with an active device within the scope of either Rule 9 or 11.</p> <p><b>Examples:</b> hemodialyzers; devices to remove white blood cells from whole blood.</p> <p><b>NOTE:</b> For the purpose of this part of the rule, ‘modification’ does not include simple, mechanical filtration or centrifuging which are covered below.</p>
<p><b>unless</b> the treatment consists of filtration, centrifuging or exchanges of gas or of heat, in which case they are in Class B.</p>	<p><b>Examples:</b> devices to remove carbon dioxide; particulate filters in an extracorporeal circulation system.</p>
<p><b>Rule 4.</b> All other non-invasive devices are in Class A.</p>	<p>These devices either do not touch the patient or contact intact skin only.</p> <p><b>Examples:</b> urine collection bottles; compression hosiery; non-invasive electrodes, hospital beds.</p>


## 6.2. Invasive Devices

Rule	Illustrative Examples
<p><b>Rule 5.</b> All invasive devices with respect to body orifices (other than those which are surgically invasive) and which:</p> <ul style="list-style-type: none"> <li>• are not intended for connection to an active medical device, or</li> <li>• are intended for connection to a Class A medical device only.</li> </ul>	<p>Such devices are invasive in body orifices and are not surgically invasive. Devices tend to be diagnostic and therapeutic instruments used in ENT, ophthalmology, dentistry, proctology, urology and gynaecology. Classification depends on the duration of use and the sensitivity (or vulnerability) of the orifice to such invasion.</p>
<p>- are in Class A if they are intended for transient use;</p>	<p><b>Examples:</b> examination gloves; enema devices</p>
<p>- are in Class B if they are intended for short-term use;</p>	<p><b>Examples:</b> urinary catheters, tracheal tubes.</p>
<p><b>unless</b> they are intended for short-term use in the oral cavity as far as the pharynx, in an ear canal up to the ear drum or in a</p>	<p><b>Examples:</b> dressings for nose bleeds</p>


 <b>Botswana Medicines Regulatory Authority</b>	Page 13 of 20
	<b>Document type:</b> Guideline
	<b>Title:</b> Guideline for Medical Devices Classification
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023

nasal cavity, in which case they are in Class A	
- are in Class C if they are intended for long-term use;	<u>Example:</u> urethral stent; contact lenses for long term continuous use (for this device, removal of the lens for cleaning is considered as part of the continuous use).
<b>unless</b> they are intended for long-term use in the oral cavity as far as the pharynx, in an ear canal up to the eardrum or in a nasal cavity and are not liable to be absorbed by the mucous membrane, in which case they are in Class B	<u>Examples:</u> orthodontic materials, removable dental prosthesis.
All invasive devices with respect to body orifices (other than those which are surgically invasive) that are intended to be connected to an active medical device in Class B or a higher class, are in Class B.	<u>Examples:</u> tracheal tubes connected to a ventilator; suction catheters for stomach drainage; dental aspirator tips. <b>NOTE:</b> Independent of the time for which they are invasive.
<b>Rule 6.</b> All surgically invasive devices intended for <b>transient use</b> are in Class B,	A majority of such devices fall into several major groups: those that create a conduit through the skin (e.g. syringe needles; lancets), surgical instruments (e.g. single-use scalpels; surgical staplers; single-use aortic punch); surgical gloves; and various types of catheter/sucker etc.
<b>unless</b> they are reusable surgical instruments, in which case they are in Class A; or	<u>Examples:</u> Manually operated surgical drill bits and saws. <b>NOTE:</b> A surgical instrument connected to an active device is in a higher class than A.
<b>unless</b> intended to supply energy in the form of ionizing radiation, in which case they are in Class C; or	<u>Example:</u> catheter containing sealed radioisotopes.
<b>unless</b> intended to have a biological effect or be wholly or mainly absorbed, in which case they are in Class C; or	<b>NOTES:</b> (a) The ‘biological effect’ referred to is an intended one rather than unintentional. The term ‘absorption’ refers to the degradation of a material within the body and the metabolic elimination of the resulting degradation products from the body. (b) This part of the rule does not apply to those substances that are excreted without modification from the body. <u>Example:</u> Insufflation gases for the abdominal cavity.


**This document is property of the Botswana Medicines Regulatory Authority (BOMRA). It is strictly confidential and may on no account be reproduced, copied or divulged to any third party without prior authorization by BOMRA Management.**

 <b>Botswana Medicines Regulatory Authority</b>	Page 14 of 20
	<b>Document type:</b> Guideline
	<b>Title:</b> Guideline for Medical Devices Classification
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023

<p><b>unless</b> intended to administer medicinal products by means of a delivery system, if this is done in a manner that is potentially hazardous taking account of the mode of application, in which they are in Class C; or</p>	<p>Example: insulin pen for self-administration.</p> <p><b>NOTE:</b> The term ‘administration of medicines’ implies storage and/or influencing the rate/volume of medicine delivered not just channelling. The term ‘potentially hazardous manner’ refers to the characteristics of the device and not the competence of the user.</p>
<p><b>unless</b> they are intended specifically for use in direct contact with the central nervous system, in which case they are in Class D; or</p>	<p>Example: spinal needle.</p>
<p><b>unless</b> intended specifically to diagnose, monitor or correct a defect of the heart or of the central circulatory system through direct contact with these parts of the body, in which case they are in Class D</p>	<p>Examples: angioplasty balloon catheters and related guide wires; dedicated disposable cardiovascular surgical instruments.</p>
<p><b>Rule 7.</b> All surgically invasive devices intended for short-term use are in Class B,</p>	<p>Such devices are mostly used in the context of surgery or post-operative care, or are infusion devices, or are catheters of various types.</p> <p><b>Examples:</b> infusion cannulae; temporary filling materials; non-absorbable skin closure devices; tissue stabilisers used in cardiac surgery.</p> <p><b>NOTE:</b> Includes devices that are used during cardiac surgery but do not monitor or correct a defect.</p>
<p><b>unless</b> they are intended to administer medicinal products, in which case they are in Class C; or</p>	<p><b>NOTE:</b> The term ‘administration of medicines’ implies storage and/or influencing the rate/volume of medicine delivered not just channelling.</p>
<p><b>unless</b> they are intended to undergo chemical change in the body (except if the devices are placed in the teeth), in which case they are in Class C; or</p>	<p>Example: surgical adhesive.</p>
<p><b>unless</b> they are intended to supply energy in the form of ionizing radiation, in which case they are in Class C; or</p>	<p>Example: brachytherapy device.</p>

 <b>Botswana Medicines Regulatory Authority</b>	Page 15 of 20
	<b>Document type:</b> Guideline
	<b>Title:</b> Guideline for Medical Devices Classification
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023


<p><b>unless</b> they are intended to have a biological effect or to be wholly or mainly absorbed, in which case they are in Class D; or</p>	<p><u>Example:</u> absorbable suture; biological adhesive.</p> <p><b>NOTE:</b> The ‘biological effect’ referred to is an intended one rather than unintentional. The term ‘absorption’ refers to the degradation of a material within the body and the metabolic elimination of the resulting degradation products from the body.</p>
<p><b>unless</b> they are intended specifically for use in direct contact with the central nervous system, in which case they are in Class D;</p>	<p><u>Example:</u> neurological catheter.</p>
<p><b>unless</b> they are intended specifically to diagnose, monitor or correct a defect of the heart or of the central circulatory system through direct contact with these parts of the body, in which case they are in Class D</p>	<p><u>Examples:</u> cardiovascular catheters; temporary pacemaker leads; carotid artery shunts.</p>
<p><b>Rule 8.</b> All implantable devices, and long-term surgically invasive devices, are in Class C,</p>	<p>Most of the devices covered by this rule are implants used in the orthopaedic, dental, ophthalmic, and cardiovascular fields. <u>Example:</u> maxilla-facial implants; bone plates and screws; bone cement; non-absorbable internal sutures; posts to secure teeth to the mandibula bone (without a bioactive coating).</p>
<p><b>unless</b> they are intended to be placed into the teeth or on prepared tooth structure, in which case they are in Class B; or</p>	<p><u>Examples:</u> materials for inlays, crowns, and bridges; dental filling materials.</p>
<p><b>unless</b> they are intended to be used in direct contact with the heart, the central circulatory system or the central nervous system, in which case they are in Class D; or</p>	<p><u>Examples:</u> prosthetic heart valves; cardiovascular stents; pacemaker leads and electrodes; deep brain stimulation electrodes; cerebrospinal catheter.</p>
<p><b>unless</b> they are intended to be life supporting or life sustaining, in which case they are in Class D; or</p>	

 <b>Botswana Medicines Regulatory Authority</b>	Page 16 of 20
	<b>Document type:</b> Guideline
	<b>Title: Guideline for Medical Devices Classification</b>
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023

<b>unless</b> they are intended to be active implantable medical devices, in which case they are Class D; or	<u>Example:</u> pacemakers; implantable defibrillators.
<b>unless</b> they are intended to have a biological effect or to be wholly or mainly absorbed, in which case they are in Class D; or	<u>Example:</u> implants claimed to be bioactive. <b>NOTE:</b> Hydroxy-apatite is considered as having biological effect only if so claimed and demonstrated by the manufacturer.
<b>unless</b> they are intended to administer medicinal products, in which case they are in Class D; or	<u>Example:</u> subcutaneous infusion ports for long term use.
<b>unless</b> they are intended to undergo chemical change in the body (except if the devices are placed in the teeth), in which case they are in Class D; or	<u>Example:</u> surgical adhesives intended for long term use. <b>NOTE:</b> Bone cement is not within the scope of the term 'chemical change in the body' since any change takes place in the short rather than long term.
<b>unless</b> they are breast implants, in which case they are in Class D.	


### 6.3 Active Devices

Rule	Illustrative Examples
<b>Rule 9(i).</b> All active therapeutic devices intended to administer or exchange energy are in Class B,	Such devices are mostly electrically powered equipment used in surgery; devices for specialised treatment and some stimulators. <u>Examples:</u> muscle stimulators; powered dental hand pieces; hearing aids; neonatal phototherapy equipment; ultrasound equipment for physiotherapy.
<b>unless</b> their characteristics are such that they may administer or exchange energy to or from the human body in a potentially hazardous way, including ionizing radiation, taking account of the nature, the density and site of application of the energy, in which case they are in Class C.	<u>Examples:</u> lung ventilators; baby incubators; electrosurgical generators; external pacemakers and defibrillators; surgical lasers; lithotriptors; therapeutic X-ray and other sources of ionizing radiation. <b>NOTE:</b> The term 'potentially hazardous' refers to the type of technology involved and the intended application.

 <b>Botswana Medicines Regulatory Authority</b>	Page 17 of 20
	<b>Document type:</b> Guideline
	<b>Title:</b> Guideline for Medical Devices Classification
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023

<b>Rule 9(ii).</b> All active devices intended to control or monitor the performance of active therapeutic devices in Class C, or intended directly to influence the performance of such devices, are in Class C.	<u>Examples:</u> external feedback systems for active therapeutic devices.
<b>Rule 10(i).</b> Active devices intended for diagnosis are in Class B:	Such devices include equipment for ultrasonic diagnosis/imaging, capture of physiological signals.
- if they are intended to supply energy which will be absorbed by the human body (except for devices used solely to illuminate the patient's body, with light in the visible or near infra-red spectrum, in which case they are Class A), or	<u>Examples:</u> magnetic resonance equipment; diagnostic ultrasound in non-critical applications; evoked response stimulators.
- if they are intended to image in vivo distribution of radiopharmaceuticals, or	<u>Example:</u> gamma/nuclear cameras
- if they are intended to allow direct diagnosis or monitoring of vital physiological processes,	<u>Example:</u> electronic thermometers, stethoscopes and blood pressure monitors; electrocardiographs
<b>unless</b> they are specifically intended for: a) monitoring of vital physiological parameters, where the nature of variations is such that it could result in immediate danger to the patient, for instance variations in cardiac performance, respiration, activity of central nervous system, or	<u>Example:</u> monitors/alarms for intensive care; biological sensors; oxygen saturation monitors; apnoea monitors.
b) diagnosing in clinical situations where the patient is in immediate danger, in which case they are in Class C.	<u>Example:</u> ultrasound equipment for use in interventional cardiac procedures.
<b>Rule 10(ii).</b> Active devices intended to emit ionizing radiation and intended for diagnostic and/or interventional radiology, including devices which control or monitor such devices, or those which directly influence their performance, are in Class C.	<u>Example:</u> devices for the control, monitoring or influencing of the emission of ionizing radiation.


**This document is property of the Botswana Medicines Regulatory Authority (BOMRA). It is strictly confidential and may on no account be reproduced, copied or divulged to any third party without prior authorization by BOMRA Management.**

 <b>Botswana Medicines Regulatory Authority</b>	Page 18 of 20
	<b>Document type:</b> Guideline
	<b>Title: Guideline for Medical Devices Classification</b>
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023


<p><b>Rule 11.</b> All active devices intended to administer and/or remove medicinal products, body liquids or other substances to or from the body are in Class B,</p>	<p>Such devices are mostly drug delivery systems or anaesthesia equipment.</p> <p><u>Examples:</u> suction equipment; feeding pumps; jet injectors for vaccination; nebuliser to be used on conscious and spontaneously breathing patients where failure to deliver the appropriate dosage characteristics is not potentially hazardous</p>
<p><b>unless</b> this is done in a manner that is potentially hazardous, taking account of the nature of the substances involved, of the part of the body concerned and of the mode and route of administration, in which case they are in Class C.</p>	<p><u>Examples:</u> infusion pumps; anaesthesia equipment; dialysis equipment; hyperbaric chambers; nebuliser where the failure to deliver the appropriate dosage characteristics could be hazardous.</p>
<p><b>Rule 12.</b> All other active devices are in Class A.</p>	<p><u>Examples:</u> examination lamps; surgical microscopes; powered hospital beds &amp; wheelchairs; powered equipment for the recording, processing, viewing of diagnostic images; dental curing lights.</p>

#### 6.4 Additional Rules

Rule	Illustrative Examples
<p><b>Rule 13.</b> All devices incorporating, as an integral part, a substance which, if used separately, can be considered to be a registered therapeutic/ medicinal product, and which is liable to act on the human body with action ancillary to that of the devices, are in Class D.</p>	<p>These medical devices incorporate medicinal substances in an ancillary role.</p> <p><u>Examples:</u> antibiotic bone cements; heparin coated catheters; wound dressings incorporating antimicrobial agents to provide ancillary action on the wound; blood bags incorporating an anti-coagulant.</p> <p><b>NOTE:</b> “Integral part” means that the device and the therapeutic/ medicinal substance are physically or chemically combined at the time of administration (i.e. use, implantation, application etc.) to the patient</p> <p><b>NOTE:</b> This rule does not apply to products that incorporated therapeutic/ medicinal substances that are not intended to act on the human body e.g. catheter surface modification or</p>

 <b>Botswana Medicines Regulatory Authority</b>	Page 19 of 20
	<b>Document type:</b> Guideline
	<b>Title:</b> Guideline for Medical Devices Classification
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023

	coating with silver or heparin substance that is solely intended to protect the catheter.
<b>Rule 14.</b> All devices manufactured from or incorporating animal or human cells/tissues/derivatives thereof, whether viable or non-viable, are in Class D,	<u>Example:</u> porcine heart valves, catgut sutures
<b>unless</b> such devices are manufactured from or incorporate non-viable animal tissues or their derivatives that come in contact with intact skin only in which case they are in Class A.	<u>Examples:</u> leather components of orthopaedic appliances.
<b>Rule 15.</b> All devices intended specifically to be used for sterilising or disinfecting medical devices as the end point of processing are in Class B.	<u>Example:</u> desk-top sterilisers for use with dental instruments. Disinfectants intended to be used with medical devices
<b>unless</b> they are disinfectant solutions or washer-disinfectors intended specifically for invasive medical devices, as the end point of processing, in which case they are in Class C; or	<u>Examples:</u> solutions intended to be used for the disinfection of medical devices without further processing (for example in a steriliser) including those where the infective agent is a prion; washer-disinfectors specifically for disinfecting an endoscope or another invasive device.
<b>unless</b> they are intended to clean medical devices by means of physical action only, in which case they are in Class A.	
<b>Rule 16.</b> All devices that are intended specifically to be used for disinfecting, cleaning, rinsing or, when appropriate, hydrating contact lenses are in Class C.	
<b>Rule 17.</b> All devices used for contraception or the prevention of the transmission of sexually transmitted diseases are in Class C,	<u>Examples:</u> condoms; contraceptive diaphragms
<b>unless</b> they are implantable or long-term invasive devices, in which case they are in Class D.	<u>Example:</u> intrauterine contraceptive device.

 <b>Botswana Medicines Regulatory Authority</b>	Page 20 of 20
	<b>Document type:</b> Guideline
	<b>Title:</b> Guideline for Medical Devices Classification
<b>Function:</b> Medical Devices	<b>Document No:</b> BOMRA/ER/MED/P04/G05
<b>Department:</b> Product Evaluation and Registration	<b>Issue No:</b> 1.0
	<b>Effective date:</b> 01-04-2023

## 7. Rationale for inclusion of the Additional Rules

There are a small number of products that fall within the scope of the definition of a medical device and which may need to be classified to take account of factors other than those covered by the general rules (Rules 1 to 12). While the authority continues to support and encourage regulatory harmonisation, the Authority may reflect different local needs or social considerations when it introduces regulations on the classification of a minority of medical devices. Therefore, five Additional Rules are provided (Rules 13 to 17).

Matters that may need to be considered are: -

Rule 13: Devices incorporating a medicinal substance

- i. The regulations applying to medicinal products require different acceptance procedures to those for medical devices. •
- ii. The behavior of a medicinal substance used in conjunction with a medical device may differ from that covered by its approved use as a medicinal product alone.

Rule 14: Devices incorporating animal or human tissues

- i. There is an absence of global regulatory controls for such devices.
- ii. Classification needs to acknowledge the diversity of opinions on such devices, globally
- iii. The possible transmission of infectious agents to human beings by the use of devices incorporating animal or human tissues (e.g. Bovine Spongiform Encephalopathies (BSE) and Creutzfeldt-Jacob disease (CJD)) demands classification at a higher level

Rule 15: Disinfection as the end point of processing

- i. Classification of disinfection solutions and washer-disinfector equipment intended for the treatment of invasive devices as the end point of processing rather than as an intermediate step before sterilisation.

Rule 16: Fluids used with contact lenses

- i. The particular concerns relating to disinfectant solutions and other fluids that are used with contact lenses, due to sensitivity and vulnerability of the eye.

Rule 17: Contraceptive devices

- i. The hazard associated with unwanted pregnancy if caused by mechanical failure of the device.
- ii. The need to safeguard public health through the use of condoms to reduce the prevalence of sexually transmitted diseases.
- iii. User expectation that contraceptive devices are perfectly reliable and safe despite published data to the contrary.